



# EMSL Analytical, Inc.

200 Route 130 North, Cinnaminson, NJ 08077

Phone/Fax: (800) 220-3675 / (856) 786-0327

<http://www.EMSL.com>

[cinnaminsonradonlab@emsl.com](mailto:cinnaminsonradonlab@emsl.com)

EMSL Order: 382300083

CustomerID: INDS42

CustomerPO:

ProjectID:

Attn: **Elessar Connolly-Fairchild**  
**Indy Safety**  
**5010 NE Roselawn St**  
**Portland, OR 97218**

Phone: (503) 501-7396  
Fax:  
Received: 1/6/2023 10:40 AM  
Analysis Date: 1/6/2023  
Collected: 1/3/2023

Project: **Banks School District Retest**

## Banks School District Retest

### Test Report: Radon in Air Test Results

Liquid Scintillation ID	Location	Radon Activity (pCi/L)	Start	Stop	Temperature F	Humidity %	Sample Type
517871 382300083-0001	Coachs Office Boys Locker	1.5	1/3/2023 1:37:00 PM	1/5/2023 1:56:00 PM	67	45	Customer
<b>Sample Notes:</b>							
520574 382300083-0002	Blank	0	1/3/2023 1:37:00 PM	1/5/2023 1:56:00 PM	67	45	Blank
<b>Sample Notes:</b>							

#### Report Notes

The United States Environmental Protection Agency (EPA) has established a radon action level of 4.0pCi/L. EPA recommends mitigation of a structure if the confirmed radon level is equal to or greater than 4.0pCi/L. The radon tests were performed using liquid scintillation radon detectors and counted on a liquid scintillation counter following EPA Method # 402-R-92-004 testing protocol for Radon in Air testing. EPA recommends retesting every two years. Please contact EMSL Analytical, Inc. or your State Health Department for further information. All procedures used for generating this report are in complete accordance with the current EPA protocols for the analysis of Radon in Air.

Analyst(s)

*Symone Tibbs (2)*

Dominic Gehret, Radiochemistry Laboratory Manager, NJ  
Radon Measurement Specialist MES 13910  
or other approved signatory

In no event shall EMSL be liable for indirect, special, consequential, or incidental damages, including, but not limited to, damages for loss of profit or goodwill regardless of the negligence (either sole or concurrent) of EMSL and whether EMSL has been informed of the possibility of such damages, arising out of or in connection with EMSL's services thereunder or the delivery, use, reliance upon or interpretation of test results by client or any third party. We accept no legal responsibility for the purposes for which the client uses the test results. In no event shall EMSL be liable to a client or any third party, whether based upon theories of tort, contract or any other legal or equitable theory, in excess of the amount paid to EMSL by client thereunder. The test results meets all NELAC requirements unless otherwise specified.

Samples analyzed by EMSL Analytical, Inc. Cinnaminson, NJ FL RB2034/R2687,IL RNL2008202,IN RTL00935,IA RNLAB10005,KS KS-LB-0005/KS-MS-0482,ME SPC202,MN RL-0005,NE 474/RMB-1083,NJ 03036/MEB92525/MES13910,NY 10872,OH RL39,OK D9952,PA 2573/3393/68-00367,RI RMB-108/RI00179,WV RL000220,NRSB-ARL6006,NRPP

Initial report from 01/09/2023 15:39:00